



Kalibracijski certifikat

Calibration Certificate

Številka certifikata

Certificate number

Naročnik kalibracije

Customer

Mesto kalibracije

Site of Calibration

Predmet:

Object

Proizvajalec:

Manufactures:

Tip merila:

Type:

Identifikacijska številka:

Identification number:

Nosilnost

Capacity

d

Readability

Datum kalibracije

Date of calibration

Sledljivost

Traceability

Kalibracijski certifikat dokumentira sledljivost do nacionalnih etalonov, ki realizirajo merske enote v skladu z Mednarodnim sistemom merskih enot (SI).

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of units (SI).

Izjave

Declarations

Dovoljeno je razmnoževanja celotnega certifikata. Razmnoževanje posameznih delov je dovoljeno samo s pisno odobritvijo laboratorija. Brez podpisa in žiga kalibracijski certifikat ni veljaven.

Kopijo certifikata hranimo najmanj 5 let.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced with written approval of the laboratory. Calibration certificate without signature and seal are not valid.

A copy of this calibration certificate will be kept for a period of at least 5 years.

Kalibracijski postopek in merilni rezultati z merilno negotovostjo

Calibration procedure and measurement results with uncertainty of measurement

(glej stran 2)

(look page 2)

Odobril

Approved by

Vodja oziroma namestnik vodje Laboratorija za kalibracije

Head or Deputy of Head of Calibration Laboratory

Datum izdaje

Date of issue

Kalibracijski postopek

Calibration procedure

Številka certifikata

Certificate number

Kalibracija neavtomatske tehtnice je bila izvedena v skladu z internim kalibracijskim postopkom AM07N02, ki temelji na Kalibracijskem vodilu EURAMET/cg-18/v.03 "Guidelines on the Calibration of Non-Automatic Weighing Instruments", in z uporabo delovnih etalonov sledljivih do mednarodnega prototipa kilograma, ki ga hrani BIMP.

Calibration of the non-automatic weighing instrument was carried out in accordance with the laboratory's internal calibration procedure AM07N02, which is based on Calibration Guide EURAMET/cg-18/v.03 "Guidelines on the Calibration of Non-Automatic Weighing Instruments", and using working standard traceable to the international kilogram prototype kept at BIMP.

Pogoji okolja Temperatura

Environmental conditions

Temperature

Kazanje tehtnice pred naravnavanjem

Indication of the instrument before adjusting

(referenčno breme = Max m_{ref})

(reference load = Max m_{ref})

Naravnavanje instrumenta pred kalibracijo da/yes ne/no

Adjustment the instrument before calibration

avtomatsko /automatic

zunanje / external

Predobremenitev da/yes ne/no

Preloading

Kazanje tehtnice je bilo določeno z ločljivostjo:

The indication was determined with resolution:

v območju do / in the range up to

v območju do / in the range up to

v območju do / in the range up to

Merilni rezultati in merilna negotovost

Measurement results and uncertainty of measurement

m_{ref} referenčno breme reference load	I kazanje indication	T tara tare load	E pogrešek kazanja error of indication	TOL max dovoljen pogrešek MPE	K faktor pokritja coverage factor	U Razš. mer. negotov expanded measurement uncertainty
1						
2						
3						
4						
5						
6						

Pogrešek kazanja je razlika med kazanjem tehtnice in maso referenčnega bremena:

The error of indication is the difference between the indication of the instrument and the reference mass:

$$E = I - m_{ref}$$

TOL (MPE)

Največji (maksimalen) dovoljen pogrešek je naveden na željo naročnika kalibracije merila.

The largest (maximum) authorized error is indicated on the customer's request calibration criteria.

Podana razširjena merilna negotovost, ki jo dobimo, ko standardno negotovost pomnožimo s faktorjem pokritja k (podan je v zgornji preglednici) ki pri normalni porazdelitvi ustreza verjetnosti približno 95%. Merilno negotovost smo ocenili v skladu z dokumentom EA-04/02.

Vse vrednosti so vrednosti konvencionalne mase. Podane vrednosti veljajo za predmet kalibracije in pogoje okolice v času kalibracije. V izračunu ni zajeta komponenta dolgoročne nestabilnosti predmeta kalibracije.

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k (given in the table above), which for normal distribution corresponds to a coverage probability of approximately 95%. The uncertainty of measurement has been evaluated in accordance with EA-04/02.

All values are conventional mass values. The values stated apply for the condition of the balance and the environment at the time of calibration. There is no component included for long term instability of the instrument. If the indications and errors have been determined with higher resolution than by the normal resolution, the reported uncertainty is smaller than what would be found with normal readings.